





## GLOVE SPECIFICATIONS

Item Number	4030
Product Group	Disposable Gloves
Description	<b>Dura-Cor™</b> , Latex Examination Gloves, Powder Free, <b>11-mil Palm Thickness, 14-mil Fingertip Thickness</b> , 300 mm. length (12 inches), Blue Color, Beaded Cuff, Textured Finish, Low Protein Content
<b>4030</b>  <u><b>Class 1 Medical Device</b></u> <b>U.S.:</b> Complies with FDA 510(k) requirements. <b>Europe:</b> Complies with 93/42/EEC  <b>ASTM D3578:</b> Standard Specifications for Rubber Exam Gloves  <b>ASTM D5151:</b> Standard Test Method for Detection of Holes in Medical Gloves  Accepted Quality Level (AQL): 1.5  <b>ISO 9001: 2000</b> Certified Manufacturing Facility  <u><b>Food Contact Information:</b></u> <b>U.S.:</b> Component Materials comply with all federal regulations for food contact. (U.S. Code of Federal Regulations, 21CFR, Parts 170-199) <b>Europe:</b> Complies with 90/128/EEC	  <p><b>Caution:</b> This product contains Natural Rubber Latex, which may cause allergic reactions in some individuals.</p>
Unit of Sale	One Case (500 pieces)
Case Pack	10 x 50 pc. dispensers
Country of Origin	Sri Lanka Or Malaysia
Available Sizes	S-XL

### Corporate Headquarters:

4025 Viscount Avenue  
Memphis, TN. 38118  
Phone 901-458-5030  
Fax 901-458-3343